



Fast Selective Detection of Pyocyanin Using Cyclic Voltammetry

Alatraktchi, Fatima AlZahra'a; Andersen, Sandra Breum; Johansen, Helle Krogh; Molin, Søren; Svendsen, Winnie Edith

Published in:
Sensors

DOI:
[10.3390/s16030408](https://doi.org/10.3390/s16030408)

Publication date:
2016

Document version
Publisher's PDF, also known as Version of record

Document license:
[CC BY](#)

Citation for published version (APA):
Alatraktchi, F. A., Andersen, S. B., Johansen, H. K., Molin, S., & Svendsen, W. E. (2016). Fast Selective Detection of Pyocyanin Using Cyclic Voltammetry. *Sensors*, 16(3), [408]. <https://doi.org/10.3390/s16030408>

Article

Fast Selective Detection of Pyocyanin Using Cyclic Voltammetry

Fatima AlZahra'a Alatraktchi ^{1,2,3,*}, Sandra Breum Andersen ², Helle Krogh Johansen ^{2,4}, Søren Molin ^{2,3} and Winnie E. Svendsen ¹

¹ Department of Micro- and Nanotechnology, Technical University of Denmark, Ørstedes Plads, 2800 Kgs. Lyngby, Denmark; Winnie.Svendsen@nanotech.dtu.dk

² Novo Nordisk Foundation Center for Biosustainability, Technical University of Denmark, Kogle Allé 6, 2970 Hørsholm, Denmark; sandrabreumandersen@gmail.com (S.B.A.); hkj@biosustain.dtu.dk (H.K.J.); sm@bio.dtu.dk (S.M.)

³ Department of Systems Biology, Technical University of Denmark, Kemitorvet, 2800 Kgs. Lyngby, Denmark

⁴ Department of Clinical Microbiology, Afsnit 9301, Rigshospitalet, Juliane Maries Vej 22, 2100 København, Denmark

* Correspondence: faaat@nanotech.dtu.dk; Tel.: +45-4525-6842; Fax: +45-4588-7762

Academic Editor: W. Rudolf Seitz

Received: 8 December 2015; Accepted: 15 March 2016; Published: 19 March 2016

Abstract: Pyocyanin is a virulence factor uniquely produced by the pathogen *Pseudomonas aeruginosa*. The fast and selective detection of pyocyanin in clinical samples can reveal important information about the presence of this microorganism in patients. Electrochemical sensing of the redox-active pyocyanin is a route to directly quantify pyocyanin in real time and *in situ* in hospitals and clinics. The selective quantification of pyocyanin is, however, limited by other redox-active compounds existing in human fluids and by other metabolites produced by pathogenic bacteria. Here we present a direct selective method to detect pyocyanin in a complex electroactive environment using commercially available electrodes. It is shown that cyclic voltammetry measurements between -1.0 V to 1.0 V reveal a potential detection window of pyocyanin of 0.58 – 0.82 V that is unaffected by other redox-active interferences. The linear quantification of pyocyanin has an R^2 value of 0.991 across the clinically relevant concentration range of 2 – 100 μ M. The proposed method was tested on human saliva showing a standard deviation of $2.5\% \pm 1\%$ ($n = 5$) from the known added pyocyanin concentration to the samples. This inexpensive procedure is suggested for clinical use in monitoring the presence and state of *P. aeruginosa* infection in patients.

Keywords: electrochemical detection; pyocyanin; quorum sensing; diagnosis; cyclic voltammetry

1. Introduction

Pseudomonas aeruginosa is reported to be among the most problematic bacteria in infections such as bloodstream, surgical wound, burn wound, and cystic fibrosis lung infections [1–3]. This opportunistic pathogen is seldom responsible for infections in healthy individuals, but is successful in immune-compromised patients, where infection causes significant morbidity and mortality [4,5]. While factors causing the infections to occur are well known, little is documented about the signals that allow benign bacteria to become pathogenic [6]. Thus, information regarding the pathogenic behavior of *P. aeruginosa* can be gained by monitoring the initial signals produced by this bacterium.

Special for *P. aeruginosa* is the production of the redox-active molecule pyocyanin. Pyocyanin production is controlled by quorum sensing [7]. *P. aeruginosa* uses quorum sensing to make collective decisions about virulence expression. Pyocyanin is assumed to be released prior to virulent activity and may itself be a quorum sensing signal [8,9]. Thus, the ability to assess the pyocyanin level produced

in patients can reveal valuable information about the state of progression of the infection before a clinical infection is apparent. The redox-active nature of pyocyanin which is uniquely produced by a problematic bacterium such as *P. aeruginosa* makes it an excellent biomarker to determine whether a patient is in danger.

In the clinic, body fluids such as urine, blood and sputum are used to determine if a patient is infected with bacteria. This is done via a microbiological culture of a sample from the body fluid to verify if an infection is present. Sputum cultures are specifically used to help identify the types of infections in the lungs and airways of cystic fibrosis patients. Sputum is normally not produced by healthy individuals, but can be produced in small quantities if irritation of the airways occurs, such as in the case of smokers and asthma patients. The presence of bacteria in sputum implies that it is possible to detect pyocyanin directly in sputum samples from patients. However, this also means that pyocyanin exists in a background of complex body fluids and this needs to be taken into account when performing measurements.

Currently, the detection of pyocyanin is accomplished by high-performance liquid chromatography (HPLC) or spectrophotometry. However, these are mostly time-consuming and costly approaches due to the pre-purification of samples. They also require isolation and culturing of bacterial samples. Electrochemical sensing is an increasingly popular method for the measurement of biochemical compounds due to the ability of specific and sensitive detection of desired molecules [10]. One of the advantages of electrochemical sensors is that they can be incorporated into point-of-care (POC) devices, providing fast and real-time diagnosis of the infection state in patients without any pre-treatment [11].

It is possible to detect pyocyanin by electrochemical sensors due to its redox-active nature. There have only been a few reports on detecting pyocyanin using electrochemistry. Sharp *et al.* presented a carbon fiber sensor for the electrochemical sensing of pyocyanin capable of detecting pyocyanin concentrations between 1 μM and 100 μM [12]. Webster and Goluch were able to detect pyocyanin with palladium hydride reference electrodes integrated in a microfluidic up-concentration device [13]. More recently, Sismaet *et al.* detected pyocyanin produced by *P. aeruginosa* by biochemically up-regulating the pyocyanin production [14]. These methods are starting to make their way into the goal of clinical detection of pyocyanin in patients, some of them also by using commercially available electrodes [15]. It is common for the studies that detect pyocyanin electrochemically to find that pyocyanin is exclusively detected at negative potentials around -250 mV to -300 mV , claiming that no other chemicals can interfere with this signal [16,17]. Conversely, interferents from dead cells in human fluids can be released and detected at negative potentials, creating misleading results that could falsely be identified as pyocyanin. In reality, pyocyanin is produced in an environment of redox-active precursors and metabolites that, from an electrochemical perspective, are close to pyocyanin [18–20]. Hence, the way to clinical diagnosis of infections using pyocyanin as an infection biomarker starts with the ability to selectively detect pyocyanin among interfering compounds.

The results from this study provide a specific method to selectively detect pyocyanin in a complex mixture of interacting compounds using relatively high positive potentials. In addition, this method has been applied on human saliva samples spiked with pyocyanin. The closest body fluid to sputum is saliva, which has been used in the model of this study. The Department of Clinical Microbiology at Rigshospitalet in Copenhagen has shown that only 50% of all sputum samples are representative of the actual lower airway condition of the patients. The sputum samples were mixed with the saliva produced by the patient, which makes saliva important to characterize before any attempt of direct pyocyanin detection in sputum samples from patients [21]. With background in the results of this study, we suggest using this method for early infection diagnostics in sputum from patients with lung infections.

2. Experimental Section

2.1. Reagents

Stock solutions of 100 μM pyocyanin (P0046-25MG, SIGMA, Copenhagen, Denmark), pyoverdine (P8374-1MG, SIGMA, Copenhagen, Denmark), NADP (N5755 SIGMA, Copenhagen, Denmark), NADPH (N5130 SIGMA, Copenhagen, Denmark), NAD (N1636, SIGMA-ALDRICH, Copenhagen, Denmark), NADH (N4505, SIGMA, Copenhagen, Denmark), phenazine- $\text{C}_{12}\text{H}_8\text{N}_2$ (P13207 ALDRICH, Copenhagen, Denmark) and Lysogeny Broth (LB) medium were prepared in MilliQ water, respectively.

2.2. Preparation of Samples

Dilutions between 1 μM and 100 μM were prepared for all the compounds. A mixture of the mentioned compounds was prepared containing 100 μM of each compound including pyocyanin. This mixture will be referred to as mix-1. A corresponding mixture without addition of pyocyanin was also prepared, and referred to as mix-2. A dilution series in the range between 0 μM and 100 μM pyocyanin was prepared in MilliQ water and in background concentrations of 5, 25, 50 and 100 μM mix-2, respectively.

2.3. Cyclic Voltammetry Measurements—Experimental Protocol

Disposable screen-printed electrodes with a three-electrode configuration were used for the electrochemical experiments (C223AT, Dropsens, Spain). The electrodes consisted of a 1.6 mm gold working electrode, a gold counter electrode and a silver reference electrode. The electrodes were connected to a potentiostat (Metrohm Autolab, The Netherlands) from which cyclic voltammetry was used to characterize the electrochemical profile of the different compounds and to quantify the pyocyanin content. The software Autolab NOVA 1.10 (Metrohm Autolab, The Netherlands) was used for data handling, peak finding and analysis.

All the compounds were individually characterized by five cyclic voltammetry sweeps between -1.0 V to 1.0 V and sweep rates of 0.05 V/s using concentrations of $100\text{ }\mu\text{M}$ of each compound.

Cyclic voltammograms between -1.0 V to 1.0 V and a scan rate of 0.05 V/s were used to characterize the profiles of mix-1 and mix-2. The calibration curve was obtained by detecting the current difference for the peaks of cyclic voltammograms performed on the dilution series of pyocyanin and mix-1, respectively; using a potential window of -0.4 V to 1.0 V and a scan rate of 0.05 V/s . All the measurements were conducted *versus* the Ag reference electrode.

2.4. Cyclic Voltammetry of Human Saliva and Artificial Sputum

Human saliva (approximately 2 mL) was collected from healthy volunteers fasting 12 h prior to the experiment. The saliva was placed on the electrodes and all air bubbles were gently removed. A cyclic voltammetry scan from -1.0 V to 1.3 V was performed using a scan rate of 0.10 V/s . Saliva was mixed with $100\text{ }\mu\text{M}$ mix-2 and $160\text{ }\mu\text{M}$ pyocyanin and placed on new electrodes before measuring the current in a potential window of -1.0 V to 1.0 V using a scan rate of 0.10 V/s . This was repeated five times.

Artificial sputum (ASM) was prepared according to Kirchner *et al.* [22]. A cyclic voltammetry scan from -1.0 V to 1.0 V was performed using a scan rate of 0.10 V/s . All the measurements were conducted *versus* the Ag reference electrode.

3. Results and Discussion

3.1. The Selectivity Window of Pyocyanin

The preliminary assessment of the electrochemical profile of pyocyanin using cyclic voltammetry shows several oxidation peaks at -0.560 V , -0.311 V and 0.699 V (Figure 1, blue curve). The MilliQ

water control gave no peaks, as expected (black curve). It is likely that the peak at -0.311 V is equivalent to the pyocyanin peaks located around -0.25 V to 0.30 V reported by earlier studies [6,12,13].




Figure 1. Cyclic voltammogram of $500 \mu\text{M}$ pyocyanin (blue) compared to the MilliQ water control (black) using a scan rate of 0.05 V/s . No signals are generated from cyclic voltammetry of MilliQ water, while pyocyanin reveals a characteristic profile with oxidation peaks at -0.560 V, -0.311 V and 0.699 V (arrows) measured *versus* the Ag reference electrode.

The characteristic peak potentials of NAD, NADH, NADP, NADP, pyoverdine, phenazine, LB, human saliva and ASM were extracted from their respective cyclic voltammograms (Figure 2). The individual cyclic voltammograms of the different compounds can be viewed in the Supplementary data. The detection window of each compound was determined by the characteristic potential peak that increases proportionally with the increasing concentration. The bars represent the lower and higher potential intervals which the max peak lies within. Several compounds are represented twice as two or more peaks corresponded to the increasing concentration. Figure 2 shows that pyocyanin has two detection windows that are located outside the characteristic detection windows of the interfering compounds, namely at -0.560 V and at 0.699 V. In contrast, the window of detection with a max peak at -0.311 V overlaps with human saliva and ASM. This implies that it will not be possible to distinguish if a peak in this potential window belongs to pyocyanin or the tested body fluid in a cyclic voltammogram. The implication of this interference is the main reason why the relatively positive peak potential is more beneficial to use in the quantification experiments compared to the negative peak potential. The unique and independent detection window of pyocyanin between 0.58 V and 0.82 V allows selectivity in detection among the other redox-active chemicals.




Figure 2. Potential peak values of the different compounds and related detection width (bars) extracted from cyclic voltammograms of the individual compounds. The detection window of pyocyanin with peak at 0.699 V and corresponding start and end potentials lies outside the interaction width of the other redox-active compounds.

As the max potential peak of 0.699 V was chosen for further quantification, a suitable scan range from -0.4 V to 1.0 V was established for the following quantitative investigations. This relatively high potential has never been applied in pyocyanin quantification in previous research. Therefore, the electrochemical behavior of pyocyanin at these relatively high potentials in a mixture of redox-active compounds is not yet available for clinical use. The increase of electrode potential is a self-enhancing process. When the electrode potential is increased, the ohmic potential drops and the signal-to-noise ratio is enhanced. Thus, the use of relatively high characteristic potentials might enhance the limit of detection of pyocyanin, when considering use of, e.g., square wave voltammetry or amperometry.

3.2. Identifying Electrochemical “Fingerprint” of Pyocyanin among Interacting Compounds

A clear electrochemical fingerprint of pyocyanin was observed when detected among other redox-active compounds that were equal in concentration to pyocyanin. The cyclic voltammogram of mix-1 clearly reveals the presence of pyocyanin in the sample as a peak at 0.68 V (Figure 3, red curve). On the other hand, no pyocyanin peak is apparent in the cyclic voltammogram of mix-2 (black curve). Mix-2 gives a superposition of an oxidation peak at 0.48 V and a width that is slightly overlapping with the pyocyanin detection window. In mix-1 it seems likely that the superposition of the peaks of the interfering compounds has shifted to peak around 0.10 V, which is close to the picture given by Figure 2. Another remarkable observation is that the potential window with a peak at -0.560 V detected in the cyclic voltammogram of pure pyocyanin (Figures 1 and 2) did not appear in Figure 3. This may be due to a non-reversible oxidation of pyocyanin.




Figure 3. Cyclic voltammograms of mixture of compounds containing pyocyanin (mix-1, red curve) and mixture without pyocyanin (mix-2, black curve) measured *versus* the reference electrode. The red curve has a peak at 0.68 V confirming the presence of pyocyanin (arrow), while it is absent in the black curve of mix-2.

3.3. Detection Limit of Pyocyanin in a Complex Mixture of Interfering Compounds

After characterizing the signature of pyocyanin among the interferents, the limit of detection (LOD) was investigated. A new sensor was used for each measurement to avoid cross-contamination or fouling of electrodes. The peak currents in the detection window of pyocyanin were extracted from cyclic voltammograms of pyocyanin in different mix concentrations as presented in Figure 4. The peak currents were stable despite the increase of mix concentration. However, the LOD of pyocyanin was affected by the presence of the increasing mix concentration. The LOD was 2 μM obtained in the absence of any interfering compounds. At 5 μM mix concentration, a detection limit of 3 μM was obtained. When the mix concentration was increased to 25 μM , the detection limit increased to 6 μM and remained the same up to 100 μM mix background.




Figure 4. Detection of pyocyanin in different mix backgrounds between 5 μM and 100 μM in addition to pyocyanin detection in MilliQ water (zero mix background). The current peaks of the varying pyocyanin concentrations were unaffected by the mix concentration. Detection of 2 μM pyocyanin was only possible when no interfering signals were present. When the mix concentration was above 25 μM the detection limit of pyocyanin remained 6 μM .

The max peak current of pyocyanin as function of concentration has a linearity fit of $R^2 = 0.991$ (Figure 5). The LOD represents the lowest detectable concentration where a peak was detected by the NOVA1.11 software in the cyclic voltammograms. The LOD is 2 μM , when no mix is present in the samples. The error bars are a measure of the standard deviation of the measurements of pyocyanin in varying mix backgrounds and measurements in MilliQ water. The detection in varying mix backgrounds mimics reality better than using a constant background concentration as the background will always be unknown. The small error bars reveal that pyocyanin can be detected independently by cyclic voltammetry in its respective detection window regardless of the concentration of interfering compounds with an upper detection limit of 6 μM .




Figure 5. Detection limit of pyocyanin using max peak currents of cyclic voltammograms. The limit of detection (LOD) is 2 μM . The linear fit to peak currents *vs.* pyocyanin concentration has a linearity of 0.991. The error bars represent the standard deviation from measurements in the four different backgrounds of mix concentrations and in no mix background.

The peak potential of pyocyanin shifts slightly to higher values as a function of pyocyanin concentration (Figure 6). The fit indicates a slight increase of potential as a function of concentration. As the pyocyanin concentration decreases, the corresponding potential peak approximates the lower limit of the pyocyanin detection window. This may explain the detection limit around 2 μM .




Figure 6. Peak potential as function of pyocyanin concentration. The potential peaks go towards the lower detection window of pyocyanin as the pyocyanin concentration decreases.

3.4. Detection of Pyocyanin in Saliva Samples

An example of a cyclic voltammogram of healthy human saliva is plotted along with a cyclic voltammogram of human saliva containing 160 μM pyocyanin and 50 μM mix-2 in Figure 7. Mix-2 was added in order to complicate the background and observe if it is still possible to detect the pyocyanin. Complicating the background is important since the interference caused by dead cells in the human saliva is not necessarily equal from person to another. Peaks at -0.20 V are seen in both measurements while a peak at 0.60 V only appears in the saliva sample containing pyocyanin. The peak potential lies within the selectivity window of pyocyanin, thus confirming the sensing of pyocyanin. The peak current is 3.31×10^{-6} , corresponding to 164 ± 2 μM ($n = 5$) pyocyanin according to the calibration curve in Figure 5. It is a deviation of $2.5\% \pm 1\%$ SE from the known pyocyanin concentration added to the samples.




Figure 7. Example of pyocyanin detection in human saliva sample. Black illustrates a cyclic voltammogram of human saliva without pyocyanin. Dark red cyclic voltammogram is of human saliva containing pyocyanin and mix. The peak around 0.60 V versus the reference electrode confirms the presence of pyocyanin in the sample (arrow).

Earlier studies assume that no interference occurs at negative potentials [13,14,16]. However, the peaks appearing at -0.20 V in the saliva samples can easily interfere with the signals of pyocyanin if a negative detection window is used to detect pyocyanin. Once again, the potential window with a peak at -0.560 V detected in the cyclic voltammogram of pure pyocyanin (Figures 1 and 2) did not appear in the cyclic voltammogram of saliva spiked with pyocyanin in Figure 7. This leaves the optimal reliable detection window for this particular situation to be the positive potential window.

Another peak around 0.11 V is observed in both cyclic voltammograms in Figure 7. This peak might belong to mix compounds as the potential is similar to what was observed in Figure 3. Saliva comprises mainly dead cells that are normally shed from the mucosal surface of the tongue and gums and from the inside of the cheeks [19]. Dead cells release metabolites such as NAD, NADH, NADP and NADPH which might have been detected before being taken up by living bacteria. As the background interference caused by dead cells might differ from patient to patient, the proposed method is an easy way to selectively detect pyocyanin without considering the non-constant background. This is especially beneficial when measuring in sputum samples collected from children, where a significant amount of water and saliva comes up with the expectorate.

The ability to detect pyocyanin at positive potentials is advantageous both with regard to avoiding interfering signals and with regard to enhancing the signal-to-noise ratio [23]. Although the LOD was not as low as in other studies, this work is a proof of concept that it is possible to detect pyocyanin at relatively high positive potentials while preserving the selectivity and speed of sensing. In further studies, the sensors can be optimized and an electrolyte other than MilliQ water can be used to improve the electrochemical detection. As the proof of concept has been established in this work, submicromolar LOD can be investigated using alternative techniques to cyclic voltammetry. For instance, chronoamperometry is a more sensitive electrochemical quantification tool that is likely to give higher resolution and detection limits due to its high signal-to-noise ratio [24,25].

Saliva is the body fluid closest to sputum. Today sputum samples from cystic fibrosis patients are used for bacterial culturing to identify the type of infection. If the sensors can be used to detect bacteria in sputum samples, it will be possible to avoid lengthy culturing procedures and costly bacterial sequencing. This work demonstrates a direct, sensitive and specific method to determine medically relevant pyocyanin levels in complex samples from patients without any pretreatment procedures [6,26].

4. Conclusions

Commercially available, disposable, screen-printed electrodes were used to selectively detect pyocyanin among interfering redox-active compounds. For the first time, a relatively high potential window of 0.58 – 0.82 V was used to identify pyocyanin in cyclic voltammograms. The advantage of using this potential window is that it lies outside the characteristic potential windows of other interfering compounds that could exist in human samples. The concept was successfully tested on human saliva and artificial sputum, which are the closest fluids to real sputum samples from patients. The detection of biologically important pyocyanin concentrations by disposable electrodes in complex backgrounds opens up the potential application for real-time *in situ* measurements in clinical samples or during surgeries.

Acknowledgments: The authors sincerely acknowledge Arto Heiskanen, Department of Micro- and Nanotechnology at DTU Nanotech, and Jaime Castillo-Leon, Sol Voltaics AB, for valuable discussions. Special appreciation to Alex Toftgaard Nielsen, Novo Nordisk Foundation Center for Biosustainability-Technical University of Denmark, who has contributed with valuable knowledge regarding the metabolism of bacteria. Helle Krogh Johansen was funded by a clinical research stipend from The Novo Nordisk Foundation and Rigshospitalet Rammebevilling 2015-17 and Lundbeckfonden Grant R167-2013-15229.

Author Contributions: Fatima AlZahra'a Alatraktchi designed, conducted and analyzed all the experimental work in the reported research in addition to writing and reviewing the paper. Sandra Breum Andersen contributed with scientific discussion of the content and reviewing the manuscript. Helle Krogh Johansen, Winnie E. Svendsen

and Søren Molin contributed by supervising and reviewing the paper. All authors read and approved the final manuscript.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Miller, L.C.; O'Loughlin, C.T.; Zhang, Z.; Siryaporn, A.; Silpe, J.E.; Bassler, B.L.; Semmelhack, M.F. Development of potent inhibitors of pyocyanin production in *Pseudomonas aeruginosa*. *J. Med. Chem.* **2015**, *58*, 1298–1306. [[CrossRef](#)] [[PubMed](#)]
2. Damkiær, S.; Yang, L.; Molin, S.; Jelsbak, L. Evolutionary remodeling of global regulatory networks during long-term bacterial adaptation to human hosts. *Proc. Natl. Acad. Sci. USA* **2013**, *110*, 7766–7771. [[CrossRef](#)] [[PubMed](#)]
3. Abu, E.A.; Su, S.; Sallans, L.; Boissy, R.E.; Greatens, A.; Heineman, W.R.; Hassett, D.J. Cyclic voltammetric, fluorescence and biological analysis of purified aeruginosin A, a secreted red pigment of *Pseudomonas aeruginosa* PAO1. *Microbiology* **2013**, *159*, 1736–1747. [[CrossRef](#)] [[PubMed](#)]
4. Rada, B.; Gardina, P.; Myers, T.G.; Leto, T.L. Reactive oxygen species mediate inflammatory cytokine release and EGFR-dependent mucin secretion in airway epithelial cells exposed to *Pseudomonas* pyocyanin. *Mucosal. Immunol.* **2011**, *4*, 158–171. [[CrossRef](#)] [[PubMed](#)]
5. Rada, B.; Leto, T.L. Pyocyanin effects on respiratory epithelium: Relevance in *Pseudomonas aeruginosa* airway infections. *Trends Microbiol.* **2013**, *21*, 73–81. [[CrossRef](#)] [[PubMed](#)]
6. Webster, T.A.; Goluch, E.D. Electrochemical detection of pyocyanin in nanochannels with integrated palladium hydride reference electrodes. *Lab. Chip* **2012**, *12*, 5195–5201. [[CrossRef](#)] [[PubMed](#)]
7. Michel-Briand, Y.; Baysse, C. The pyocins of *Pseudomonas aeruginosa*. *Biochimie* **2002**, *84*, 499–510. [[CrossRef](#)]
8. Jayaseelan, S.; Ramaswamy, D.; Dharmaraj, S. Pyocyanin: Production, applications, challenges and new insights. *World J. Microb. Biotechnol.* **2014**, *30*, 1159–1168. [[CrossRef](#)] [[PubMed](#)]
9. Dietrich, L.E.P.; Price-Whelan, A.; Petersen, A.; Whiteley, M.; Newman, D.K. The phenazine pyocyanin is a terminal signalling factor in the quorum sensing network of *Pseudomonas aeruginosa*. *Mol. Microbiol.* **2006**, *61*, 1308–1321. [[CrossRef](#)] [[PubMed](#)]
10. Alatraktchi, F.A.; Bakmand, T.; Dimaki, M.; Svendsen, W.E. Novel Membrane-Based Electrochemical Sensor for Real-Time Bio-Applications. *Sensors* **2014**, *14*, 22128–22139. [[CrossRef](#)] [[PubMed](#)]
11. Svendsen, W.E.; Alatraktchi, F.A.; Bakmand, T.; Waagepetersen, H.; Dimaki, M. Novel culturing platform for brain slices and neuronal cells. *IEEE Eng. Med. Biol. Soc.* **2015**, 3299, 346–349.
12. Sharp, D.; Gladstone, P.; Smith, R.B.; Forsythe, S.; Davis, J. Approaching intelligent infection diagnostics: Carbon fibre sensor for delectrochemical pyocyanin detection. *Bioelectrochemistry* **2009**, *77*, 114–119. [[CrossRef](#)] [[PubMed](#)]
13. Sismaet, H.J.; Webster, T.A.; Goluch, E.D. Up-regulating pyocyanin production by amino acid addition for early electrochemical identification of *Pseudomonas aeruginosa*. *Analyst* **2014**, *139*, 4241–4246. [[CrossRef](#)] [[PubMed](#)]
14. Seviour, T.; Doyle, L.E.; Lauw, S.J.L.; Hinks, J.; Rice, S.A.; Netsatyy, V.J.; Webster, R.D.; Kjelleberg, S.; Marsili, E. Voltammetric profiling of redox-active metabolites expressed by *Pseudomonas aeruginosa* for diagnostic purposes. *Chem. Commun.* **2015**, *51*, 3789–3792. [[CrossRef](#)] [[PubMed](#)]
15. Webster, T.A.; Sismaet, H.J.; Conte, J.L.; Chan, I.C.; Goluch, E.D. Electrochemical detection of *Pseudomonas aeruginosa* in human fluid samples via pyocyanin. *Biosens. Bioelectron.* **2014**, *60*, 265–270. [[CrossRef](#)] [[PubMed](#)]
16. Bellin, D.L.; Sakhtah, H.; Rosenstein, J.K.; Levine, P.M.; Thimot, J.; Emmett, K.; Dietrich, L.E.P.; Shepard, K.L. Integrated circuit-based electrochemical sensor for spatially resolved detection of redox-active metabolites in biofilms. *Nat. Commun.* **2014**, *5*, 3256. [[CrossRef](#)] [[PubMed](#)]
17. Munteanu, G.; Dempsey, E.; McCormac, T.; Munteanu, C. Fast cyclic voltammetry of redox system NAD⁺/NADH on the copper nanodoped mercury monolayer carbon fiber electrode. *J. Electroanal. Chem.* **2012**, *665*, 12–19. [[CrossRef](#)]
18. Lin, K.-C.; Chen, S.-M. Reversible cyclic voltammetry of the NADH/NAD⁺ redox system on hybrid poly(luminol)/FAD film modified electrodes. *J. Electroanal. Chem.* **2006**, *589*, 52–59. [[CrossRef](#)]

19. Roka, A.; Inzelt, G. Cyclic voltammetric and nanogravimetric studies of NADP⁺ redox transformations on a yeast-modified platinum electrode. *Electrochem. Commun.* **2014**, *45*, 9–12.
20. Godderis, L.; Schouteden, C.; Tabish, A.; Poels, K.; Hoet, P.; Baccarelli, A.A.; van Landuyt, K. Global Methylation and Hydroxymethylation in DNA from Blood and Saliva in Healthy Volunteers. *BioMed Res. Int.* **2015**, *2015*, 845041. [[CrossRef](#)] [[PubMed](#)]
21. Johansen, H.K.; Department of Clinical Microbiology, Rigshospitalet, Copenhagen, Denmark. Repræsentativt materiale til mikrobiologisk diagnostik af nedre luftvejsinfektioner. Personal communication, 2016. (In Danish)
22. Kirchner, S.; Fothergill, J.L.; Wright, E.A.; James, C.E.; Mowat, E.; Winstanley, C. Use of Artificial Sputum Medium to Test Antibiotic Efficacy Against *Pseudomonas aeruginosa* in Conditions More Relevant to the Cystic Fibrosis Lung. *Jove-Journal of Visualized Experiments. J. Vis. Exp.* **2012**, *64*, e3857. [[PubMed](#)]
23. Alatraktchi, F.A.-Z.; Molin, S.; Johansen, H.K.; Svendsen, W.E. Nano-molar selective detection of pyocyanin using electrochemical amperometry for diagnostic purposes. *Future Med.* **2016**, submitted.
24. Bard, A.J.; Faulkner, L.R. *Electrochemical Methods—Fundamentals and Applications*; Wiley: New York, NY, USA, 2001.
25. Kissinger, P.; Heineman, W.R. *Laboratory Techniques in Electroanalytical Chemistry*, 2nd ed.; CRC Press: New York, NY, USA, 1996.
26. Caldwell, C.C.; Chen, Y.; Goetzmann, H.S.; Hao, Y.; Borchers, M.T.; Hassett, D.J.; Young, L.R.; Mavrodi, D.; Thomashow, L.; Lau, G.W. *Pseudomonas aeruginosa* Exotoxin Pyocyanin Causes Cystic Fibrosis Airway Pathogenesis. *Am. J. Pathol.* **2009**, *175*, 2473–2488. [[CrossRef](#)] [[PubMed](#)]



© 2016 by the authors; licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons by Attribution (CC-BY) license (<http://creativecommons.org/licenses/by/4.0/>).